



The Department of Environment Protection and Energy Maanshan Iron & Steel Co. Ltd


2022 47

!

2022 4 A#
2#3#4#

一、炼铁北 A#烧结机机头脱硫脱硝排口

2022 4 8 A#
82.427 mg/m³



二、能控北 2#排涝泵站

2022 4 20

(CODcr) 20 80.20 mg/L 21
 155.192 mg/L 22 52.096 mg/L COD
 72
 COD

水质常规监测因子记录

*污染源: 马鞍山钢铁股份有限公司

*监测点: 能控北3#排涝泵站

*标记类型: 调试

*子类型: 自动监测设备新装调试

*开始时间: 2022-04-19 12:20:00

结束时间: 2022-04-22 14:00:59

*监测因子: pH 化学需氧量 氨氮 流量

*备注: 厂家设备调试



四、能控北 4#排涝泵站

2022 4 18 -4 20 4#
(CODcr) 18 76.10 mg/L 19 142.57
mg/L 20 71.06 mg/L COD
72 COD

编辑常规监测因子标记记录

*污染源： 马鞍山钢铁股份有限公司

*监测点： 能控北4#排涝泵站

*标记类型： 调试

*子类型： 自动监测设备新装调试

*开始时间： 2022-04-17 15:00:00


结束时间： 2022-04-20 17:00:59

*监测因子：
 pH 化学需氧量 氨氮 流量

*备注：
厂家调试

日期	COD (mg/L)	pH	氨氮 (mg/L)	流量 (m³/h)
2022-04-18	75	8.7		
2022-04-19				
2022-04-20				

备注：厂家调试



五、冷轧北水处理排口

2022 4 19 -4 20
(CODcr) 19 70.982 mg/L 20 148.195
mg/L COD 72
COD



六、四钢轧炼钢西路排口

2022 4 27

pH

pH=5.95

PH

pH

The screenshot shows a web-based monitoring system interface. At the top, there's a navigation bar with various icons and a search bar. Below that, a table lists monitoring points. A pop-up window titled '编辑常规监测因子标记记录' (Edit Regular Monitoring Factor Marking Record) is open, showing details for a specific record.

监测因子	标记类型	子类型	开始时间	结束时间	标记是否完成	备注	填报人	填报时间
1 pH	故障						刘文静	2022-04-29 15:53:14
2 非甲烷总烃	故障						王昕	2022-04-27 16:50:02
3 烟尘 二氧化硫 氮氧化物 流量	故障						王昕	2022-04-27 17:28:05
4 pH-化学需氧量-氨氮	故障						王昕	2022-04-27 16:48:51
5 pH	故障						刘文静	2022-04-27 08:56:23
6 氨氮	质控比对						赵小敏	2022-04-26 10:28:08
7 pH-化学需氧量-氨氮	质控比对						车文海	2022-04-26 10:30:48
8 烟尘 二氧化硫 氮氧化物 流量	故障						王昕	2022-04-27 17:27:58

编辑常规监测因子标记记录

*污染源: 马鞍山钢铁股份有限公司

*监测点: 鞍钢北厂炼钢西路排口

*标记类型: 故障

*子类型: 自动监测设备故障

*开始时间: 2022-04-27 20:10:00

*结束时间: 2022-04-28 20:00:59

*监测因子: pH 化学需氧量 氨氮 流量

*备注: PH探头故障, 原值后发报程序出错, 导致无法报数, 通知厂家量科对设备进行程序安装和测试。

